
AIPHONE



A simple communication system sa medical and commercial fields.

SYSTEM FEATURES

- All call and group call
- Call log
- Powerful sound

- 20/40/60/80 call capacity
- Large variety of sub stations
- Durability and reliability



tisfying a broad range of needs in

NURSE CALL SYSTEM

Provides reliable communication and enhances the quality of care.



Individual communication with patients is possible by using the bedside handheld sub station.



COMMUNICATION SYSTEM

Used in a wide range of environments such as highway toll gates and schools.






Broadcast option is available.

Necessary communication for healthcare facilities.


< Examples of System Configuration >

NIM can be used as a nurse call system for providing calls from patient rooms and bathrooms to nurse Various system configurations offer application flexibility and sophistication.

PC connection makes it possible to check call history.



stations.




Personal communication ensures the priv

Single bed communication:


Communication system using a speaker and a microphone at each bed.



■ Examples of Equipment Application





Corridor light with buzzer can be used for notification of calls from patients requiring special attention.




acy of your patients.


■ Features








■ Using the system at the master station



Push and hold TALK to begin talking with the caller.

Release $\overline{\mbox{TALK}}$ to listen.



• When receiving two or more calls*

Push the station selector button to select the desired call.**
Then, talk with the caller by using the handset or pressing the TALK button.


** Maximum of 10 calls can be received simultaneously.

**If this step is skipped, the first-arrived call is automatically selected.

neously. Blinks in red.

Calling a single sub station

Push a call selector button to call the desired sub station. Then, talk with the caller by using the handset or pressing the TALK button.




Shared communication for multiple-occupa


Multiple-user communication:

Communication system using a ceiling speaker and a microphone for communication with the entire room.



■ Examples of Equipment Application






ncy patient rooms is possible.

■ Features



The user speaks toward the ceiling microphone and hears the reply from the ceiling speaker.






Ceiling speaker NI-LB



Ceiling microphone NI-SB

Communication is also possible by using a bedside wall jack with microphone instead of the ceiling microphone.




The user speaks toward the wall jack and hears the reply from the ceiling speaker.





Wall jack with microphone for NBR-8A

NIR-MB



For more information on master station use see Page 6.

Lighted lamps in corridor enable nurses to notice calls easily and respond to the patient quickly.




Corridor light, single bed NIR-4




NURSE CALL SYSTEM >> Bathroom call button, Sub stations

Quick response is possible in the case of emergencies in bathrooms.


Receiving a call is possible from a patient in a hospital bathroom.



■ Examples of Equipment Application



Features




Lighted lamps in corridor enable nurses to notice calls easily and respond to the patient quickly.

Corridor light, single bed NIR-4

Buzzer-equipped stations notify calls by flashing LED with buzzer sound.


General purpose sub stations for multiple purposes.


Desk mount sub station can be used for internal communication.




Flush mount outdoor station allows communication from outside the hospital.

communication from outside Features









Useful for tracing call history and improving

Connecting a computer to the master station helps to manage patient information, call history, and call-response data.




■ Examples of Equipment Application



service quality.

■ Features



Example of graph display





Number of calls from sub stations

This allows nursing staff to see which patients have the highest number of calls, enabling management to schedule and prioritize care.

Number of calls during different times of day

This enables verification of peak-time trends, enabling management to adjust staffing accordingly.

Response time for calls


View the number of seconds from call to response, which can be useful in improving the level of service.

A communication system specialized for sim


< Examples of System Configuration >

NIM system can be used for reliable communication. It can also be used in challenging conditions and noisy environments. This opens the door for various business and administrative applications.

Example of Installation at a Toll Gate




Other Examples of Installation (NI-JA)




ple "call" and "communication" functions.

■ Examples of Equipment Application



■ Features




NIM system can be used for communication

< Examples of System Configuration >


NIM system can be utilized to broadcast announcements to selected stations or all stations.

Example of Installation in a School



and broadcasting.


■ Examples of Equipment Application



■ Features


Broadcast to all stations


Broadcasting to all sub stations is possible.




Group broadcasts to selected stations

Broadcasts limited to selected classrooms are possible. Push the station selector button(s) for the target sub station(s).





A pretone is heard at all the sub stations before the broadcast.



Group broadcasts can be made to up to 5 sub stations.

UPGRADING FROM NEM

Upgrade from an existing NEM system.

Handheld sub station


Original system (NEM system) Privacy is difficult to maintain using ceiling sub station.

NIM system



- The handheld sub stations are equipped with a microphone and speaker, enabling communication with individual patients.
- Private communication helps to provide more detailed care of patients.

Desk mount sub station




Note: When upgrading from NEM system to NIM system, all system components and wiring need to be changed.

Flush mount sub station

Original system (NEM system)

The previous sub station had limited application scope.





• The new sub station is vandalism- and weather- resistant. This allows installation in a wider range of areas.







IK07 (ruggedized against vandalism)

Corridor light with nameplates

Original system (NEM system)

No identification of individual calls.





NIM system

• Easy identification of calls, thanks to our new corridor lamp with nameplate.



Call log software

Original system (NEM system)


Simple communication.



NIM system


• Connectable to PC for call history, patient information, and call-response data.


No need for an Internet connection. Using a RS-232C cable is enough for PC connection.





SYSTEM CONFIGURATION

System configuration example. Wiring method is subject to system configuration.









The number of sub stations is subject to the type of master station.

Master station		Sub station (Oty)	Personal computer (Oty)
Model	Qty	Sub station (Qty)	Personal computer (Qty)
NIM-20B	1	20	1
NIM-40B	1	40	1
Master station, 60-call(*)	1	60	1
Master station, 80-call(*)	1	80	1

(*): Special order product

■ Wiring distance

	Wire diameter	Maximum distance
Master station - sub station	Ø0.65 mm (22 AWG) Ø0.9 mm (19 AWG)	130 m (490') 200 m (650')
Master station - power supply	Ø0.9 mm (19 AWG)	10 m (33')
Master station - call extension speaker (IER-2)	Ø0.9 mm (19 AWG)	50 m (165')





NIM-20B Master station, 20-call



Power source: 48 V DC (supplied from the power supply)

Communication: Handset: Voice-actuated communication

Open voice: Push-to-talk (PTT)

Material: Panel: Flame resisting resin

Chassis: SPCC

* 60 or 80 call model is also available (special order product).

Call button, wall jack



NBR-8A Handheld call button

Material: Flame resisting



NIR-1 Wall jack for NBR-8A

Power source: Supplied from the master station

Mounting: Wall (flush)-mount Electrical box: 1-gang box Material: Plastic



NIR-MB Wall jack with microphone for NBR-8A

Power source: Supplied from the master station

Mounting: Wall (flush)-mount Electrical box: 1-gang box Material: Plastic



NIR-7BS Call latching wall jack for NBR-8A

Power source: Supplied from the master station

Mounting: Wall (flush)-mount Electrical box: 1-gang box Material: Plastic



NIR-6 Call button

Power source: Supplied from the

master station
Mounting: Wall (flush)-mount
Electrical box: 1-gang box Material: Plastic

Sub station



NI-RC

Bedside handheld sub station

Power source: Supplied from the master station **Communication:** Open voice hands-free auto-voice actuation Material: Plastic



NIR-HP

Wall jack for NI-RC

Power source: Supplied from the master station Mounting: Wall (flush)-mount Electrical box: 1-gang box
Material: Plastic



NIR-HPF

Wall jack with hanger for NI-RC Power source: Supplied from the master station Mounting: Wall (flush)-mount Electrical box: 2-gang box Material: Plastic

Ceiling speaker



NI-LB

Ceiling speaker Power source: Supplied from the

master station

Mounting: Flush-mount on the ceiling Material: Plastic

Ceiling microphone



NI-SB

Ceiling microphone

Power source: Supplied from the

master station

Mounting: Flush-mount on the ceiling

Material: Stainless steel



NI-BA

Sub station, surface mount Power source: Supplied from the

master station Communication: Open voice hands-free

communication Mounting: Desktop use or wall (surface)-mount Electrical box: 1-gang box Material: Flame resisting resin



NI-JA

Sub station, flush mount Power source: Supplied from the master

station

Communication: Open voice hands-free

communication

Mounting: Flush-mount

Electrical box: 2-gang box Material: Front panel: Stainless steel
Main unit: Flame resisting

resin

Remarks: Weather-resistant (Protection class: IP54,IK07)

Corridor light, reset button



NIR-4

Corridor light, single bed

NIR-4S

Corridor light, multiple beds

Power source: Supplied from the master

station
Mounting:Wall (flush)-mount
Electrical box: 1-gang box
Material: Plastic



NIR-2

Call reset button Power source: Supplied from the master station

Mounting: Wall (flush)-mount Electrical box: 1-gang box
Material: Plastic



NIR-4BZ

Corridor light with buzzer Power source: Supplied from the master station

Mounting: Wall (flush)-mount Electrical box: 2-gang box Material: Plastic



NIR-42

Corridor light with reset button Power source: Supplied from the master station

Mounting: Wall (flush)-mount Electrical box: 2-gang box Material: Plastic



NIR-31

Corridor light with nameplate, 1 name



NIR-32

Corridor light with nameplates, 2 names



NIR-34

Corridor light with nameplates,

Power source: Supplied from the master station **Material:** SPCC



NIR-7W

Bathroom call button Power source: Supplied from the master station

Mounting: Wall (flush)-mount Electrical box: 1-gang box Material: Flame resisting resin Remarks: Splash proof

NIR-7HW

Bathroom call button with pull cord Power source: Supplied from the master station

Mounting: Wall (flush)-mount Electrical box: 1-gang box Material: Flame resisting resin Remarks: Splash proof

Call extension speaker



IER-2

Call extension speaker

Power supply



Power source: AC 100 - 240 V, 50/60 Hz. Power consumption: 120 W

(Secondary) Output voltage: DC 48 V (Secondary) Output current: 2.1 A

Mounting: Stationary, wall-mount (IS-RACK(*)) or rack-mount (IS-RACK(*))

Material: Steel plate

(*): IS-RACK: Consult an authorized Aiphone dealer for details.



NI-SOFT Call log software Accessory : RS-232C cable

PC application System requirements fo PC

OS	Windows 7 Service Pack1 / Home Premium / Professional / Ultimate(32/64bit) Windows 8.1/ Pro (32/64bit)	
Processor	3 GHz or higher	
System memory (RAM)	2 GB RAM or higher	
Hard drive	100 MB or more (Additional space is needed during installation.)	
Requirements	RS232C port installed on the PC or an add-in board CD-R readable drive A PC operable 24 hours a day Display: 1024 (M) x 768 (H) or higher, 16-bit color or higher Font: Arial Run-time: Microsoft.NET Framework 4.5	


The application may not run or may not operate correctly (such as freezes occurring) on a PC with specifications below the necessary requirements.


> 66.7 ø105

•We recommend using a PC that exceeds the basic system

Microsoft and Windows are either registered trademarks or trademarks of Microsoft Corporation in the United States and/or other countries.

UNIT DIMENSIONS (mm)





Approx. 30

13


Approx. 25

ø180

70

23

116



12

Approx.

70

Distributed by:



Deming Medal



JQA-0291 Corresponding product: Standard product Order product



JQA-EM0453

: HEADQUARTERS
TOYOTA FACTORY
TAIHOH SITE
DEVELOPMENT CENTER



http://www.aiphone.net/